

## Product Classification List

Your Search for

Manufacturer/Distributor:

HCPCS Code: L5615

Product Name:

Product Model:

Classification(s):

Product Name	Manufacturer/Distributor	Model Number	HCPCS Code	Effective Begin Date	Effective End Date	Comments
ALLUX	PROTEOR USA	NE-Z4SH	L5615 + L5856 + L5848 + L5845	01/01/2024		SH AT THE END OF THE MODEL NUMBER INDICATED SCREW HEAD
ALLUX 2	PROTEOR USA	NE-Z41SH	L5615 + L5856 + L5848 + L5845	01/01/2024		SH AT THE END OF THE MODEL NUMBER INDICATED SCREW HEAD
ALLUX 2	PROTEOR USA	NE-Z41	L5615 + L5856 + L5848 + L5845	01/01/2024		
ALLUX	PROTEOR USA	NE-Z4	L5615 + L5856 + L5848 + L5845	01/01/2024		



October 17, 2020

CRAIG ARMSTRONG  
PROTEOR USA  
1236 W SOUTHERN AVE #101  
TEMPE, AZ 85282

**DCN Number:20244003000002**

Manufacturer Name	Product Name	Model Number	Assigned HCPCS Code(s)
PROTEOR USA	ALLUX 2	NE-Z41	L5613+L5856+L5848 +L5845
PROTEOR USA	ALLUX 2	NE-Z41SH	L5613+L5856+L5848 +L5845

Dear CRAIG ARMSTRONG,

The Pricing, Data Analysis, and Coding (PDAC) Contractor has reviewed the product(s) listed above and has approved the listed Healthcare Common Procedure Coding System (HCPCS) code(s) for billing the four Durable Medical Equipment Medicare Administrative Contractors (DME MACs).

The PDAC Contractor provides coding assistance to manufacturers to ensure proper coding of Durable Medical Equipment, Prosthetics, Orthotics, and Supplies (DMEPOS). The PDAC publishes coding decisions based on the coding guidelines established by the Local Coverage Determinations (LCDs) and associated Policy Articles and any related Advisory Articles established by the DME MACs. All products submitted to the PDAC for a coding verification review are examined by coders and professionals following a formal, standardized process.

Based on this review and application of DME MAC policy, the HCPCS code(s) listed below should be used when billing the DME MACs:

If you disagree with this decision, you may request a reconsideration within 45 days of the letter's



date and provide evidence to substantiate a reconsideration of PDAC's original coding determination. To request a reconsideration, complete the Reconsideration Request form located on the PDAC website at [www.dmepdac.com](http://www.dmepdac.com). If your request for a reconsideration is made after the 45-day time frame, it will require a new application and documentation to support the request.

It is the responsibility of manufacturers and distributors to notify the PDAC immediately of any changes involving their products, as listed on the Product Classification List (PCL) on the Durable Medical Equipment Coding System (DMECS). Further information for requesting updates to the PCL can be found on the PDAC website at [www.dmepdac.com](http://www.dmepdac.com). It is also the responsibility of manufacturers and distributors to assure their websites and product marketing materials accurately reflect the product reviewed by the PDAC and the coding decision assigned.

An assignment of the HCPCS code(s) to product(s) is not an approval or endorsement of the product(s) by Medicare or Palmetto GBA; nor does it imply or guarantee claim reimbursement or coverage.

If you have questions, please contact the PDAC HCPCS Helpline at (877) 735-1326 during the hours of 9:30 a.m. to 5:00 p.m. ET, Monday through Friday. You may also visit our [website](#) to chat with one of our representatives or select the Contact Us button at the top of the page for email, FAX or postal mail information.

Sincerely,

Pricing, Data Analysis, and Coding Contract (PDAC)  
Palmetto GBA, LLC  
[www.dmepdac.com](http://www.dmepdac.com)